



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Initial and Continuing Calibration Verification

Client: ENTACT

SDG No.: Q1282

Project: 540 Degraw St, Brooklyn, NY - E9309

RunNo.: LB134549

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV Hexavalent Chromium	mg/L	0.500	0.5	100	90-110	02/04/2025
Sample ID: CCV1 Hexavalent Chromium	mg/L	0.497	0.5	99	90-110	02/04/2025
Sample ID: CCV2 Hexavalent Chromium	mg/L	0.503	0.5	101	90-110	02/04/2025

Initial and Continuing Calibration Verification

Client: ENTACT

SDG No.: Q1282

Project: 540 Degraw St, Brooklyn, NY - E9309

RunNo.: LB134556

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV1						
Bromide	mg/L	10.3	10	103	90-110	01/20/2025
Chloride	mg/L	3.1	3	103	90-110	01/20/2025
Fluoride	mg/L	2.1	2	105	90-110	01/20/2025
Nitrite	mg/L	3.1	3	103	90-110	01/20/2025
Nitrate	mg/L	2.6	2.5	104	90-110	01/20/2025
Sulfate	mg/L	15.4	15	103	90-110	01/20/2025
Orthophosphate as P	mg/L	5.2	5	104	90-110	01/20/2025
Sample ID: CCV1						
Bromide	mg/L	10.3	10	103	90-110	02/04/2025
Chloride	mg/L	3.1	3	103	90-110	02/04/2025
Fluoride	mg/L	2	2	100	90-110	02/04/2025
Nitrite	mg/L	3.1	3	103	90-110	02/04/2025
Nitrate	mg/L	2.6	2.5	104	90-110	02/04/2025
Sulfate	mg/L	15.2	15	101	90-110	02/04/2025
Orthophosphate as P	mg/L	4.7	5	94	90-110	02/04/2025
Sample ID: CCV2						
Bromide	mg/L	10.3	10	103	90-110	02/04/2025
Chloride	mg/L	3.1	3	103	90-110	02/04/2025
Fluoride	mg/L	2	2	100	90-110	02/04/2025
Nitrite	mg/L	3.1	3	103	90-110	02/04/2025
Nitrate	mg/L	2.6	2.5	104	90-110	02/04/2025
Sulfate	mg/L	15.3	15	102	90-110	02/04/2025
Orthophosphate as P	mg/L	5.2	5	104	90-110	02/04/2025

Initial and Continuing Calibration Verification

Client: ENTACT

SDG No.: Q1282

Project: 540 Degraw St, Brooklyn, NY - E9309

RunNo.: LB134573

Analyte		Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID:	ICV						
Flash Point		o F	81.7	81	101	78-84	02/05/2025

Initial and Continuing Calibration Verification

Client: ENTACT

SDG No.: Q1282

Project: 540 Degraw St, Brooklyn, NY - E9309

RunNo.: LB134584

Analyte		Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID:	ICV						
pH		pH	6.98	7	100	90-110	02/05/2025
Sample ID:	CCV1						
pH		pH	2.01	2.00	101	90-110	02/05/2025
Sample ID:	CCV2						
pH		pH	12.02	12.00	100	90-110	02/05/2025

Initial and Continuing Calibration Verification

Client: ENTACT

SDG No.: Q1282

Project: 540 Degraw St, Brooklyn, NY - E9309

RunNo.: LB134632

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV1 TKN	mg/L	4.6	5	92	90-110	02/07/2025
Sample ID: CCV1 TKN	mg/L	4.9	5	98	90-110	02/07/2025
Sample ID: CCV2 TKN	mg/L	4.9	5	98	90-110	02/07/2025

Initial and Continuing Calibration Verification

Client: ENTACT

SDG No.: Q1282

Project: 540 Degraw St, Brooklyn, NY - E9309

RunNo.: LB134632

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
---------	-------	--------	------------	---------------	---------------------------	------------------