

Initial and Continuing Calibration Blank Summary

Client: ENTACT

SDG No.: Q1282

Project: 540 Degraw St, Brooklyn, NY - E9309

RunNo.: LB134549

Analyte	Units	Result	Acceptance Limits	Conc Qual	MDL	RDL	Analysis Date
Sample ID: ICB Hexavalent Chromium	mg/L	< 0.0050	0.0050	U	0.0027	0.01	02/04/2025
Sample ID: CCB1 Hexavalent Chromium	mg/L	< 0.0050	0.0050	U	0.0027	0.01	02/04/2025
Sample ID: CCB2 Hexavalent Chromium	mg/L	< 0.0050	0.0050	U	0.0027	0.01	02/04/2025

Initial and Continuing Calibration Blank Summary

Client: ENTACT

SDG No.: Q1282

Project: 540 Degraw St, Brooklyn, NY - E9309

RunNo.: LB134556

Analyte	Units	Result	Acceptance Limits	Conc Qual	MDL	RDL	Analysis Date
Sample ID: ICB1							
Bromide	mg/L	< 1.0000	1.0000	U	0.034	2	01/20/2025
Chloride	mg/L	< 0.3000	0.3000	U	0.011	0.6	01/20/2025
Fluoride	mg/L	< 0.2000	0.2000	U	0.057	0.4	01/20/2025
Nitrite	mg/L	< 0.3000	0.3000	U	0.011	0.6	01/20/2025
Nitrate	mg/L	< 0.2500	0.2500	U	0.0034	0.5	01/20/2025
Sulfate	mg/L	< 1.5000	1.5000	U	0.032	3	01/20/2025
Orthophosphate as P	mg/L	< 0.5000	0.5000	U	0.079	1	01/20/2025
Sample ID: CCB1							
Bromide	mg/L	< 1.0000	1.0000	U	0.034	2	02/04/2025
Chloride	mg/L	< 0.3000	0.3000	U	0.011	0.6	02/04/2025
Fluoride	mg/L	< 0.2000	0.2000	U	0.057	0.4	02/04/2025
Nitrite	mg/L	< 0.3000	0.3000	U	0.011	0.6	02/04/2025
Nitrate	mg/L	< 0.2500	0.2500	U	0.0034	0.5	02/04/2025
Sulfate	mg/L	< 1.5000	1.5000	U	0.032	3	02/04/2025
Orthophosphate as P	mg/L	< 0.5000	0.5000	U	0.079	1	02/04/2025
Sample ID: CCB2							
Bromide	mg/L	< 1.0000	1.0000	U	0.034	2	02/04/2025
Chloride	mg/L	< 0.3000	0.3000	U	0.011	0.6	02/04/2025
Fluoride	mg/L	< 0.2000	0.2000	U	0.057	0.4	02/04/2025
Nitrite	mg/L	< 0.3000	0.3000	U	0.011	0.6	02/04/2025
Nitrate	mg/L	< 0.2500	0.2500	U	0.0034	0.5	02/04/2025
Sulfate	mg/L	< 1.5000	1.5000	U	0.032	3	02/04/2025
Orthophosphate as P	mg/L	< 0.5000	0.5000	U	0.079	1	02/04/2025

Initial and Continuing Calibration Blank Summary

Client: ENTACT

SDG No.: Q1282

Project: 540 Degraw St, Brooklyn, NY - E9309

RunNo.: LB134632

Analyte	Units	Result	Acceptance Limits	Conc Qual	MDL	RDL	Analysis Date
Sample ID: ICB1 TKN	mg/L	< 0.2500	0.2500	U	0.18	0.5	02/07/2025
Sample ID: CCB1 TKN	mg/L	< 0.2500	0.2500	U	0.18	0.5	02/07/2025
Sample ID: CCB2 TKN	mg/L	< 0.2500	0.2500	U	0.18	0.5	02/07/2025

Initial and Continuing Calibration Blank Summary

Client: ENTACT

SDG No.: Q1282

Project: 540 Degraw St, Brooklyn, NY - E9309

RunNo.: LB134632

Analyte	Units	Result	Acceptance Limits	Conc Qual	MDL	RDL	Analysis Date
---------	-------	--------	----------------------	--------------	-----	-----	------------------