

Duplicate Sample Summary

Client:	ENTACT	SDG No.:	Q1282
Project:	540 Degraw St, Brooklyn, NY - E9309	Sample ID:	Q1252-01
Client ID:	EFFLUENTDUP	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
TSS	mg/L	+/-5	1630		1630		1	0.25		02/05/2025

Duplicate Sample Summary

Client:	ENTACT	SDG No.:	Q1282
Project:	540 Degraw St, Brooklyn, NY - E9309	Sample ID:	Q1271-04
Client ID:	3190DUP	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
Flash Point	o F	+/-2	>212.0		>212.0		1	0		02/05/2025

Duplicate Sample Summary

Client: ENTACT	SDG No.: Q1282
Project: 540 Degraw St, Brooklyn, NY - E9309	Sample ID: Q1282-01
Client ID: SW-WTS-02MSD	Percent Solids for Spike Sample: 0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/AD	Qual	Analysis Date
Bromide	mg/L	+/-20	12.2		12.2		1	0		02/04/2025
Fluoride	mg/L	+/-20	2.70		2.70		1	0		02/04/2025
Nitrite	mg/L	+/-20	7.00		7.00		1	0		02/04/2025
Orthophosphate as P	mg/L	+/-20	5.40		5.40		1	0		02/04/2025
Chloride	mg/L	+/-20	52.0	OR	52.0	OR	1	0		02/04/2025
Nitrate	mg/L	+/-20	17.2	OR	17.2	OR	1	0		02/04/2025
Sulfate	mg/L	+/-20	965	OR	966	OR	1	0		02/04/2025

Duplicate Sample Summary

Client: ENTACT Project: 540 Degraw St, Brooklyn, NY - E9309 Client ID: 0131-CDUP	SDG No.: Q1282 Sample ID: Q1298-07 Percent Solids for Spike Sample: 0
---	--

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
pH	pH	+/-20	6.99		7.00		1	0.14		02/05/2025

Duplicate Sample Summary

Client: ENTACT Project: 540 Degraw St, Brooklyn, NY - E9309 Client ID: COMPDUP	SDG No.: Q1282 Sample ID: Q1300-02 Percent Solids for Spike Sample: 0
---	--

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
BOD5	mg/L	+/-20	3240		3170		1	2.06		02/05/2025