

Laboratory Control Sample Summary

Client: ENTACT

SDG No.: Q1282

Project: 540 Degraw St, Brooklyn, NY - E9309

Run No.: LB134549

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	lb134549BS							
Hexavalent Chromium	mg/L	0.5	0.51		101	1	90-111	02/04/2025

Laboratory Control Sample Summary

Client: ENTACT

SDG No.: Q1282

Project: 540 Degraw St, Brooklyn, NY - E9309

Run No.: LB134556

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB134556BSW							
Bromide	mg/L	10	10.4		104	1	90-110	02/04/2025
Chloride	mg/L	3	3.10		103	1	90-110	02/04/2025
Fluoride	mg/L	2	2.00		100	1	90-110	02/04/2025
Nitrite	mg/L	3	3.10		103	1	90-110	02/04/2025
Nitrate	mg/L	2.5	2.60		104	1	90-110	02/04/2025
Sulfate	mg/L	15	15.4		103	1	90-110	02/04/2025
Orthophosphate as P	mg/L	5	5.10		102	1	90-110	02/04/2025

Laboratory Control Sample Summary

Client: ENTACT

SDG No.: Q1282

Project: 540 Degraw St, Brooklyn, NY - E9309

Run No.: LB134570

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB134570BS							
TSS	mg/L	550	532		97	1	90-110	02/05/2025

Laboratory Control Sample Summary

Client: ENTACT

SDG No.: Q1282

Project: 540 Degraw St, Brooklyn, NY - E9309

Run No.: LB134586

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB134586BS							
BOD5	mg/L	198	190		96	1	84.6-115.4	02/05/2025

Laboratory Control Sample Summary

Client: ENTACT

SDG No.: Q1282

Project: 540 Degraw St, Brooklyn, NY - E9309

Run No.: LB134632

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	PB166559BS							
TKN	mg/L	5	4.90		98	1	90-110	02/07/2025