



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 2b -

CRDL STANDARD FOR AA & ICP

Client: ENTACT **SDG No.:** Q1282
Contract: ENTA05 **Lab Code:** CHEM **Case No.:** Q1282 **SAS No.:** Q1282
Initial Calibration Source: _____
Continuing Calibration Source: _____

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CRA	Mercury	0.20	0.2	99	40 - 160	CV	02/06/2025	12:01	LB134599
CRI01	Cadmium	5.87	6.0	98	40 - 160	P	02/07/2025	12:35	LB134636
	Copper	21.0	20.0	105	40 - 160	P	02/07/2025	12:35	LB134636
	Lead	10.9	12.0	91	40 - 160	P	02/07/2025	12:35	LB134636
	Nickel	39.7	40.0	99	40 - 160	P	02/07/2025	12:35	LB134636
	Zinc	48.6	40.0	122	40 - 160	P	02/07/2025	12:35	LB134636