



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client:	ENTACT				SDG No.:	Q1282				
Contract:	ENTA05	Lab Code:	CHEM		Case No.:	Q1282	SAS No.:	Q1282		
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number	
ICB88	Mercury	0.20	+/-0.20	U	0.20	CV	02/06/2025	11:55	LB134599	

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: <u>ENTACT</u>		SDG No.: <u>Q1282</u>							
Contract: <u>ENTA05</u>	Lab Code: <u>CHEM</u>	Case No.: <u>Q1282</u>	SAS No.: <u>Q1282</u>						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB30	Mercury	0.20	+/-0.20	U	0.20	CV	02/06/2025	11:59	LB134599
CCB31	Mercury	0.20	+/-0.20	U	0.20	CV	02/06/2025	12:30	LB134599
CCB32	Mercury	0.20	+/-0.20	U	0.20	CV	02/06/2025	13:01	LB134599

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: ENTACT **SDG No.:** Q1282
Contract: ENTA05 **Lab Code:** CHEM **Case No.:** Q1282 **SAS No.:** Q1282

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB01	Cadmium	6.00	+/-6.00	U	6.00	P	02/07/2025	12:31	LB134636
	Copper	20.0	+/-20.0	U	20.0	P	02/07/2025	12:31	LB134636
	Lead	12.0	+/-12.0	U	12.0	P	02/07/2025	12:31	LB134636
	Nickel	40.0	+/-40.0	U	40.0	P	02/07/2025	12:31	LB134636
	Zinc	40.0	+/-40.0	U	40.0	P	02/07/2025	12:31	LB134636

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: ENTACT		SDG No.: Q1282							
Contract: ENTA05	Lab Code: CHEM	Case No.: Q1282	SAS No.: Q1282						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB01	Cadmium	6.00	+/-6.00	U	6.00	P	02/07/2025	13:01	LB134636
	Copper	20.0	+/-20.0	U	20.0	P	02/07/2025	13:01	LB134636
	Lead	12.0	+/-12.0	U	12.0	P	02/07/2025	13:01	LB134636
	Nickel	40.0	+/-40.0	U	40.0	P	02/07/2025	13:01	LB134636
	Zinc	40.0	+/-40.0	U	40.0	P	02/07/2025	13:01	LB134636
CCB02	Cadmium	0.41	+/-6.00	J	6.00	P	02/07/2025	13:53	LB134636
	Copper	20.0	+/-20.0	U	20.0	P	02/07/2025	13:53	LB134636
	Lead	12.0	+/-12.0	U	12.0	P	02/07/2025	13:53	LB134636
	Nickel	40.0	+/-40.0	U	40.0	P	02/07/2025	13:53	LB134636
	Zinc	40.0	+/-40.0	U	40.0	P	02/07/2025	13:53	LB134636
CCB03	Cadmium	0.42	+/-6.00	J	6.00	P	02/07/2025	14:45	LB134636
	Copper	20.0	+/-20.0	U	20.0	P	02/07/2025	14:45	LB134636
	Lead	12.0	+/-12.0	U	12.0	P	02/07/2025	14:45	LB134636
	Nickel	40.0	+/-40.0	U	40.0	P	02/07/2025	14:45	LB134636
	Zinc	40.0	+/-40.0	U	40.0	P	02/07/2025	14:45	LB134636
CCB04	Cadmium	0.26	+/-6.00	J	6.00	P	02/07/2025	15:37	LB134636
	Copper	20.0	+/-20.0	U	20.0	P	02/07/2025	15:37	LB134636
	Lead	12.0	+/-12.0	U	12.0	P	02/07/2025	15:37	LB134636
	Nickel	40.0	+/-40.0	U	40.0	P	02/07/2025	15:37	LB134636
	Zinc	40.0	+/-40.0	U	40.0	P	02/07/2025	15:37	LB134636
CCB05	Cadmium	6.00	+/-6.00	U	6.00	P	02/07/2025	16:28	LB134636
	Copper	20.0	+/-20.0	U	20.0	P	02/07/2025	16:28	LB134636
	Lead	12.0	+/-12.0	U	12.0	P	02/07/2025	16:28	LB134636
	Nickel	40.0	+/-40.0	U	40.0	P	02/07/2025	16:28	LB134636
	Zinc	40.0	+/-40.0	U	40.0	P	02/07/2025	16:28	LB134636
CCB06	Cadmium	0.28	+/-6.00	J	6.00	P	02/07/2025	17:20	LB134636
	Copper	20.0	+/-20.0	U	20.0	P	02/07/2025	17:20	LB134636
	Lead	12.0	+/-12.0	U	12.0	P	02/07/2025	17:20	LB134636
	Nickel	40.0	+/-40.0	U	40.0	P	02/07/2025	17:20	LB134636
	Zinc	40.0	+/-40.0	U	40.0	P	02/07/2025	17:20	LB134636
CCB07	Cadmium	6.00	+/-6.00	U	6.00	P	02/07/2025	18:10	LB134636
	Copper	20.0	+/-20.0	U	20.0	P	02/07/2025	18:10	LB134636
	Lead	12.0	+/-12.0	U	12.0	P	02/07/2025	18:10	LB134636
	Nickel	40.0	+/-40.0	U	40.0	P	02/07/2025	18:10	LB134636
	Zinc	40.0	+/-40.0	U	40.0	P	02/07/2025	18:10	LB134636
CCB08	Cadmium	0.27	+/-6.00	J	6.00	P	02/07/2025	18:27	LB134636
	Copper	20.0	+/-20.0	U	20.0	P	02/07/2025	18:27	LB134636
	Lead	12.0	+/-12.0	U	12.0	P	02/07/2025	18:27	LB134636
	Nickel	40.0	+/-40.0	U	40.0	P	02/07/2025	18:27	LB134636

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client:	<u>ENTACT</u>				SDG No.:	<u>Q1282</u>			
Contract:	<u>ENTA05</u>	Lab Code:	<u>CHEM</u>		Case No.:	<u>Q1282</u>	SAS No.:	<u>Q1282</u>	
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB08	Zinc	40.0	+/-40.0	U	40.0	P	02/07/2025	18:27	LB134636