

Metals

- 4 -

INTERFERENCE CHECK SAMPLE

Client: ENTACT **SDG No.:** Q1282
Contract: ENTA05 **Lab Code:** CHEM **Case No.:** Q1282 **SAS No.:** Q1282
ICS Source: EPA **Instrument ID:** P4

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSA01	Cadmium	2.97	1.0	297	-5	7	02/07/2025	12:40	LB134636
	Copper	9.83	2.0	492	-18	22	02/07/2025	12:40	LB134636
	Lead	8.02			-12	12	02/07/2025	12:40	LB134636
	Nickel	3.82	2.0	191	-38	42	02/07/2025	12:40	LB134636
	Zinc	5.22			-40	40	02/07/2025	12:40	LB134636
ICSAB01	Cadmium	1020	972	105	826	1120	02/07/2025	12:44	LB134636
	Copper	519	511	102	434	588	02/07/2025	12:44	LB134636
	Lead	56.3	49.0	115	37	61	02/07/2025	12:44	LB134636
	Nickel	1020	954	107	810	1100	02/07/2025	12:44	LB134636
	Zinc	1070	952	112	809	1095	02/07/2025	12:44	LB134636