

**Metals**

- 6 -

**DUPLICATE SAMPLE SUMMARY**

**Client:** ENTACT **Level:** LOW **SDG No.:** Q1282  
**Contract:** ENTA05 **Lab Code:** CHEM **Case No.:** Q1282 **SAS No.:** Q1282  
**Matrix:** Water **Sample ID:** Q1201-01 **Client ID:** TAPHHA-MW12-012725-00-T2DUP  
**Percent Solids for Sample:** NA **Duplicate ID** Q1201-01DUP **Percent Solids for Spike Sample:** NA

| Analyte | Units | Acceptance<br>Limit | Sample<br>Result | C | Duplicate<br>Result | C | RPD | Qual | M  |
|---------|-------|---------------------|------------------|---|---------------------|---|-----|------|----|
| Mercury | ug/L  | 20                  | 0.20             | U | 0.20                | U |     |      | CV |

**Metals**

- 6 -

**DUPLICATE SAMPLE SUMMARY**

**Client:** ENTACT **Level:** LOW **SDG No.:** Q1282  
**Contract:** ENTA05 **Lab Code:** CHEM **Case No.:** Q1282 **SAS No.:** Q1282  
**Matrix:** Water **Sample ID:** Q1201-01MS **Client ID:** TAPHHA-MW12-012725-00-T2MSD  
**Percent Solids for Sample:** NA **Duplicate ID** Q1201-01MSD **Percent Solids for Spike Sample:** NA

| Analyte | Units | Acceptance Limit | Sample Result | C | Duplicate Result | C | RPD | Qual | M  |
|---------|-------|------------------|---------------|---|------------------|---|-----|------|----|
| Mercury | ug/L  | 20               | 3.28          |   | 3.13             |   | 5   |      | CV |

**Metals**

- 6 -

**DUPLICATE SAMPLE SUMMARY**

**Client:** ENTACT **Level:** LOW **SDG No.:** Q1282  
**Contract:** ENTA05 **Lab Code:** CHEM **Case No.:** Q1282 **SAS No.:** Q1282  
**Matrix:** Water **Sample ID:** Q1282-01 **Client ID:** SW-WTS-02DUP  
**Percent Solids for Sample:** NA **Duplicate ID** Q1282-01DUP **Percent Solids for Spike Sample:** NA

| Analyte | Units | Acceptance Limit | Sample Result | C | Duplicate Result | C | RPD | Qual | M |
|---------|-------|------------------|---------------|---|------------------|---|-----|------|---|
| Cadmium | ug/L  | 20               | 3.00          | U | 3.00             | U |     |      | P |
| Copper  | ug/L  | 20               | 10.0          | U | 10.0             | U |     |      | P |
| Lead    | ug/L  | 20               | 6.00          | U | 6.00             | U |     |      | P |
| Nickel  | ug/L  | 20               | 2.51          | J | 2.57             | J | 2   |      | P |
| Zinc    | ug/L  | 20               | 19.7          | J | 19.4             | J | 2   |      | P |

# Metals

- 6 -

## DUPLICATE SAMPLE SUMMARY

**Client:** ENTACT **Level:** LOW **SDG No.:** Q1282  
**Contract:** ENTA05 **Lab Code:** CHEM **Case No.:** Q1282 **SAS No.:** Q1282  
**Matrix:** Water **Sample ID:** Q1282-01MS **Client ID:** SW-WTS-02MSD  
**Percent Solids for Sample:** NA **Duplicate ID** Q1282-01MSD **Percent Solids for Spike Sample:** NA

| Analyte | Units | Acceptance Limit | Sample Result | C | Duplicate Result | C | RPD | Qual | M |
|---------|-------|------------------|---------------|---|------------------|---|-----|------|---|
| Cadmium | ug/L  | 20               | 96.1          |   | 96.1             |   | 0   |      | P |
| Copper  | ug/L  | 20               | 153           |   | 152              |   | 1   |      | P |
| Lead    | ug/L  | 20               | 470           |   | 471              |   | 0   |      | P |
| Nickel  | ug/L  | 20               | 249           |   | 249              |   | 0   |      | P |
| Zinc    | ug/L  | 20               | 125           |   | 124              |   | 1   |      | P |