

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: ENTACT **SDG No.:** Q1282
Contract: ENTA05 **Lab Code:** CHEM **Case No.:** Q1282 **SAS No.:** Q1282

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB166595BS							
Cadmium	ug/L	100	97.0		97	80 - 120	P
Copper	ug/L	150	159		106	80 - 120	P
Lead	ug/L	500	482		96	80 - 120	P
Nickel	ug/L	250	249		100	80 - 120	P
Zinc	ug/L	100	104		104	80 - 120	P

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: ENTACT **SDG No.:** Q1282
Contract: ENTA05 **Lab Code:** CHEM **Case No.:** Q1282 **SAS No.:** Q1282

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB166596BS Mercury	ug/L	4.0	3.81		95	82 - 119	CV