

CALIBRATION VERIFICATION SUMMARY

Contract: ENTA05

Lab Code: CHEM Case No.: Q1282 SAS No.: Q1282 SDG NO.: Q1282

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 01/28/2025 01/28/2025

Client Sample No.: CCAL01 Date Analyzed: 02/05/2025

Lab Sample No.: AR1660CCC500 Data File : PP069503.D Time Analyzed: 16:20

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.686	5.592	5.792	499.410	500.000	-0.1
Aroclor-1016-2	5.708	5.614	5.814	512.280	500.000	2.5
Aroclor-1016-3	5.771	5.676	5.876	513.010	500.000	2.6
Aroclor-1016-4	5.867	5.774	5.974	510.520	500.000	2.1
Aroclor-1016-5	6.161	6.067	6.267	498.990	500.000	-0.2
Aroclor-1260-1	7.279	7.187	7.387	484.780	500.000	-3.0
Aroclor-1260-2	7.532	7.441	7.641	457.710	500.000	-8.5
Aroclor-1260-3	7.891	7.800	8.000	465.670	500.000	-6.9
Aroclor-1260-4	8.116	8.025	8.225	451.980	500.000	-9.6
Aroclor-1260-5	8.436	8.347	8.547	463.900	500.000	-7.2
Decachlorobiphenyl	10.262	10.179	10.379	49.810	50.000	-0.4
Tetrachloro-m-xylene	4.534	4.438	4.638	51.490	50.000	3.0

CALIBRATION VERIFICATION SUMMARY

Contract: ENTA05

Lab Code: CHEM Case No.: Q1282 SAS No.: Q1282 SDG NO.: Q1282

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 01/28/2025 01/28/2025

Client Sample No.: CCAL01 Date Analyzed: 02/05/2025

Lab Sample No.: AR1660CCC500 Data File : PP069503.D Time Analyzed: 16:20

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.928	4.834	5.034	506.400	500.000	1.3
Aroclor-1016-2	4.947	4.853	5.053	502.160	500.000	0.4
Aroclor-1016-3	5.125	5.030	5.230	499.940	500.000	0.0
Aroclor-1016-4	5.166	5.072	5.272	499.420	500.000	-0.1
Aroclor-1016-5	5.381	5.288	5.488	495.710	500.000	-0.9
Aroclor-1260-1	6.419	6.326	6.526	479.140	500.000	-4.2
Aroclor-1260-2	6.607	6.515	6.715	472.470	500.000	-5.5
Aroclor-1260-3	6.761	6.670	6.870	442.840	500.000	-11.4
Aroclor-1260-4	7.234	7.142	7.342	465.520	500.000	-6.9
Aroclor-1260-5	7.474	7.384	7.584	463.910	500.000	-7.2
Decachlorobiphenyl	8.898	8.811	9.011	45.000	50.000	-10.0
Tetrachloro-m-xylene	3.837	3.740	3.940	50.210	50.000	0.4