

CALIBRATION VERIFICATION SUMMARY

Contract: ENTA05

Lab Code: CHEM Case No.: Q1282 SAS No.: Q1282 SDG NO.: Q1282

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 01/28/2025 01/28/2025

Client Sample No.: CCAL02 Date Analyzed: 02/05/2025

Lab Sample No.: AR1660CCC500 Data File : PP069518.D Time Analyzed: 20:45

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	5.687	5.592	5.792	503.200	500.000	0.6
Aroclor-1016-2	5.708	5.614	5.814	508.860	500.000	1.8
Aroclor-1016-3	5.771	5.676	5.876	506.580	500.000	1.3
Aroclor-1016-4	5.868	5.774	5.974	520.030	500.000	4.0
Aroclor-1016-5	6.161	6.067	6.267	496.750	500.000	-0.7
Aroclor-1260-1	7.280	7.187	7.387	485.370	500.000	-2.9
Aroclor-1260-2	7.533	7.441	7.641	454.370	500.000	-9.1
Aroclor-1260-3	7.891	7.800	8.000	461.310	500.000	-7.7
Aroclor-1260-4	8.116	8.025	8.225	447.580	500.000	-10.5
Aroclor-1260-5	8.437	8.347	8.547	458.260	500.000	-8.3
Decachlorobiphenyl	10.263	10.179	10.379	49.500	50.000	-1.0
Tetrachloro-m-xylene	4.534	4.438	4.638	51.600	50.000	3.2

CALIBRATION VERIFICATION SUMMARY

Contract: ENTA05

Lab Code: CHEM Case No.: Q1282 SAS No.: Q1282 SDG NO.: Q1282

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 01/28/2025 01/28/2025

Client Sample No.: CCAL02 Date Analyzed: 02/05/2025

Lab Sample No.: AR1660CCC500 Data File : PP069518.D Time Analyzed: 20:45

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	4.928	4.834	5.034	500.200	500.000	0.0
Aroclor-1016-2	4.947	4.853	5.053	500.840	500.000	0.2
Aroclor-1016-3	5.125	5.030	5.230	505.880	500.000	1.2
Aroclor-1016-4	5.166	5.072	5.272	501.070	500.000	0.2
Aroclor-1016-5	5.382	5.288	5.488	515.390	500.000	3.1
Aroclor-1260-1	6.420	6.326	6.526	503.680	500.000	0.7
Aroclor-1260-2	6.607	6.515	6.715	498.380	500.000	-0.3
Aroclor-1260-3	6.761	6.670	6.870	462.970	500.000	-7.4
Aroclor-1260-4	7.234	7.142	7.342	491.340	500.000	-1.7
Aroclor-1260-5	7.474	7.384	7.584	487.930	500.000	-2.4
Decachlorobiphenyl	8.898	8.811	9.011	47.000	50.000	-6.0
Tetrachloro-m-xylene	3.838	3.740	3.940	50.680	50.000	1.4