

CALIBRATION VERIFICATION SUMMARY

Contract: ENTA05

Lab Code: CHEM Case No.: Q1282 SAS No.: Q1282 SDG NO.: Q1282

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 01/28/2025 01/28/2025

Client Sample No.: CCAL03 Date Analyzed: 02/06/2025

Lab Sample No.: AR1660CCC500 Data File : PP069533.D Time Analyzed: 02:03

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.690	5.592	5.792	480.730	500.000	-3.9
Aroclor-1016-2	5.712	5.614	5.814	477.900	500.000	-4.4
Aroclor-1016-3	5.774	5.676	5.876	477.690	500.000	-4.5
Aroclor-1016-4	5.871	5.774	5.974	483.970	500.000	-3.2
Aroclor-1016-5	6.165	6.067	6.267	477.070	500.000	-4.6
Aroclor-1260-1	7.283	7.187	7.387	461.580	500.000	-7.7
Aroclor-1260-2	7.537	7.441	7.641	426.980	500.000	-14.6
Aroclor-1260-3	7.895	7.800	8.000	436.790	500.000	-12.6
Aroclor-1260-4	8.120	8.025	8.225	436.250	500.000	-12.8
Aroclor-1260-5	8.440	8.347	8.547	445.230	500.000	-11.0
Decachlorobiphenyl	10.268	10.179	10.379	48.530	50.000	-2.9
Tetrachloro-m-xylene	4.537	4.438	4.638	51.150	50.000	2.3

CALIBRATION VERIFICATION SUMMARY

Contract: ENTA05

Lab Code: CHEM Case No.: Q1282 SAS No.: Q1282 SDG NO.: Q1282

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 01/28/2025 01/28/2025

Client Sample No.: CCAL03 Date Analyzed: 02/06/2025

Lab Sample No.: AR1660CCC500 Data File : PP069533.D Time Analyzed: 02:03

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	4.930	4.834	5.034	506.650	500.000	1.3
Aroclor-1016-2	4.949	4.853	5.053	503.830	500.000	0.8
Aroclor-1016-3	5.126	5.030	5.230	513.740	500.000	2.7
Aroclor-1016-4	5.168	5.072	5.272	510.090	500.000	2.0
Aroclor-1016-5	5.383	5.288	5.488	494.720	500.000	-1.1
Aroclor-1260-1	6.421	6.326	6.526	511.330	500.000	2.3
Aroclor-1260-2	6.609	6.515	6.715	504.860	500.000	1.0
Aroclor-1260-3	6.763	6.670	6.870	466.940	500.000	-6.6
Aroclor-1260-4	7.236	7.142	7.342	509.000	500.000	1.8
Aroclor-1260-5	7.476	7.384	7.584	510.110	500.000	2.0
Decachlorobiphenyl	8.900	8.811	9.011	49.440	50.000	-1.1
Tetrachloro-m-xylene	3.838	3.740	3.940	50.880	50.000	1.8