

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q1282

Client: ENTACT

Analytical Method: **8082A** **Datafile :** PP069517.D

Lab Sample ID	Parameter	Spike	Result	Units	Rec	RPD	Qual	RPD	Limits		RPD
								Qual	Low	High	
PB166566BSD	AR1016	5	4.90	ug/L	98	4			40 (61)	140 (112)	20 (20)
	AR1260	5	4.40	ug/L	88	0			40 (66)	140 (113)	20 (20)