

6C

SEMIVOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: CHEMTECH Contract: ENTA05
Lab Code: CHEM Case No.: Q1282 SAS No.: Q1282 SDG No.: Q1282
Instrument ID: BNA_F Calibration Date(s): 02/06/2025 02/06/2025
Calibration Time(s): 11:07 14:14

LAB FILE ID:		RRF2.5 = BF141472.D		RRF005 = BF141473.D		RRF010 = BF141474.D		
		RRF020 = BF141475.D		RRF040 = BF141476.D		RRF050 = BF141477.D		
COMPOUND	RRF2.5	RRF005	RRF010	RRF020	RRF040	RRF050	RRF	% RSD
2-Fluorophenol		1.235	1.290	1.365	1.289	1.273	1.276	3.8
Phenol-d6		1.619	1.629	1.704	1.613	1.599	1.612	3.1
Phenol		1.693	1.681	1.817	1.685	1.659	1.678	4.4
1,4-Dichlorobenzene		1.546	1.538	1.571	1.471	1.465	1.487	4.6
Nitrobenzene-d5		0.380	0.379	0.396	0.385	0.378	0.383	1.6
1,2,4-Trichlorobenzene		0.327	0.321	0.331	0.317	0.312	0.320	2.1
Naphthalene		1.067	1.041	1.096	1.037	1.012	1.041	3.0
2-Fluorobiphenyl		1.373	1.360	1.385	1.269	1.241	1.298	5.6
2,4,6-Tribromophenol		0.201	0.200	0.210	0.200	0.199	0.201	2.1
Terphenyl-d14		1.129	1.104	1.211	1.198	1.209	1.185	4.1