

Surrogate Summary

SW-846

SDG No.: Q1282

Client: ENTACT

Analytical Method: 8270E

Lab Sample ID	Client ID	Parameter	Spike (PPM)	Result (PPM)	Recovery (%)	Qual	Limits (%)	
							Low	High
PB166567BL	PB166567BL	2-Fluorophenol	150	129	86		15 (10)	110 (139)
		Phenol-d6	150	127	85		15 (10)	110 (134)
		Nitrobenzene-d5	100	87.1	87		30 (49)	130 (133)
		2-Fluorobiphenyl	100	88.2	88		30 (52)	130 (132)
		2,4,6-Tribromophenol	150	129	86		15 (44)	110 (137)
		Terphenyl-d14	100	79.3	79		30 (48)	130 (125)
PB166567BS	PB166567BS	2-Fluorophenol	150	138	92		15 (10)	110 (139)
		Phenol-d6	150	137	92		15 (10)	110 (134)
		Nitrobenzene-d5	100	95.5	95		30 (49)	130 (133)
		2-Fluorobiphenyl	100	95.3	95		30 (52)	130 (132)
		2,4,6-Tribromophenol	150	143	96		15 (44)	110 (137)
		Terphenyl-d14	100	100	100		30 (48)	130 (125)
PB166567BSD	PB166567BSD	2-Fluorophenol	150	137	91		15 (10)	110 (139)
		Phenol-d6	150	135	90		15 (10)	110 (134)
		Nitrobenzene-d5	100	92.5	92		30 (49)	130 (133)
		2-Fluorobiphenyl	100	92.8	93		30 (52)	130 (132)
		2,4,6-Tribromophenol	150	138	92		15 (44)	110 (137)
		Terphenyl-d14	100	98.3	98		30 (48)	130 (125)
Q1282-01	SW-WTS-02	2-Fluorophenol	150	57.2	38		15 (10)	110 (139)
		Phenol-d6	150	32.4	22		15 (10)	110 (134)
		Nitrobenzene-d5	100	100.0	100		30 (49)	130 (133)
		2-Fluorobiphenyl	100	99.9	100		30 (52)	130 (132)
		2,4,6-Tribromophenol	150	139	93		15 (44)	110 (137)
		Terphenyl-d14	100	103	103		30 (48)	130 (125)