

Surrogate Summary

SDG No.: Q1282

Client: ENTACT

Analytical Method: SW8260-Low

Lab Sample ID	Client ID	Parameter	Spike	Result	RecoveryQual	Limits	
						Low	High
Q1282-01	SW-WTS-02	1,2-Dichloroethane-d4	50	46.2	92	70 (74)	130 (125)
		Dibromofluoromethane	50	48.0	96	70 (75)	130 (124)
		Toluene-d8	50	48.5	97	70 (86)	130 (113)
		4-Bromofluorobenzene	50	44.0	88	70 (77)	130 (121)
VN0204WBL01	VN0204WBL01	1,2-Dichloroethane-d4	50	49.9	100	70 (74)	130 (125)
		Dibromofluoromethane	50	50.3	101	70 (75)	130 (124)
		Toluene-d8	50	49.2	98	70 (86)	130 (113)
		4-Bromofluorobenzene	50	44.2	88	70 (77)	130 (121)
VN0204WBS02	VN0204WBS02	1,2-Dichloroethane-d4	50	45.2	90	70 (74)	130 (125)
		Dibromofluoromethane	50	48.7	97	70 (75)	130 (124)
		Toluene-d8	50	49.8	100	70 (86)	130 (113)
		4-Bromofluorobenzene	50	47.8	96	70 (77)	130 (121)
VN0204WBSD0	VN0204WBSD02	1,2-Dichloroethane-d4	50	42.7	85	70 (74)	130 (125)
		Dibromofluoromethane	50	52.2	104	70 (75)	130 (124)
		Toluene-d8	50	48.0	96	70 (86)	130 (113)
		4-Bromofluorobenzene	50	47.9	96	70 (77)	130 (121)

