

**Laboratory Control Sample/Laboratory Control Sample Duplicate Summary**

**SW-846**

**SDG No.:** Q1282

**Client:** ENTACT

**Analytical Method:** SW8260-Low

**Datafile :** VN085655.D

| Lab Sample ID | Parameter               | Spike | Result | Unit | Rec | RPD | Qual | Low     | Limits<br>High | RPD     |
|---------------|-------------------------|-------|--------|------|-----|-----|------|---------|----------------|---------|
| VN0204WBSD02  | Methyl tert-butyl Ether | 20    | 21.8   | ug/L | 109 | 7   |      | 70 (78) | 130 (114)      | 20 (20) |
|               | Carbon Tetrachloride    | 20    | 21.3   | ug/L | 106 | 10  |      | 70 (77) | 130 (113)      | 20 (20) |
|               | Chloroform              | 20    | 19.7   | ug/L | 99  | 2   |      | 70 (79) | 130 (113)      | 20 (20) |
|               | 1,1,1-Trichloroethane   | 20    | 20.1   | ug/L | 101 | 1   |      | 70 (80) | 130 (108)      | 20 (20) |
|               | Benzene                 | 20    | 22.2   | ug/L | 111 | 7   |      | 70 (82) | 130 (109)      | 20 (20) |
|               | Toluene                 | 20    | 21.5   | ug/L | 108 | 1   |      | 70 (82) | 130 (110)      | 20 (20) |
|               | Tetrachloroethene       | 20    | 20.5   | ug/L | 103 | 3   |      | 70 (67) | 130 (123)      | 20 (20) |
|               | Ethyl Benzene           | 20    | 21.4   | ug/L | 107 | 4   |      | 70 (83) | 130 (109)      | 20 (20) |
|               | m/p-Xylenes             | 40    | 45.4   | ug/L | 114 | 4   |      | 70 (82) | 130 (110)      | 20 (20) |
|               | o-Xylene                | 20    | 23.0   | ug/L | 115 | 8   |      | 70 (83) | 130 (109)      | 20 (20) |