

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: ENTA05
 Lab Code: CHEM Case No.: Q1282 SAS No.: Q1282 SDG NO.: Q1282
 Lab File ID: VN085645.D Date Analyzed: 02/04/2025
 Instrument ID: MSVOA_N Time Analyzed: 11:05
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	157058	13.788				
UPPER LIMIT	314116	14.288				
LOWER LIMIT	78529	13.288				
EPA SAMPLE NO.						
SW-WTS-02	125407	13.79				
VN0204WBL01	132287	13.79				
VN0204WBS02	150317	13.79				
VN0204WBSD02	163365	13.79				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.