


Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021125\
Data File : VX044892.D
Acq On : 11 Feb 2025 08:50
Operator : JC/MD
Sample : BFB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X021025W.M
Title : SW846 8260
Last Update : Tue Feb 11 03:41:08 2025



AutoFind: Scans 1637, 1638, 1639; Background Corrected with Scan 1631

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	20.3	9120	PASS
75	95	30	60	53.4	24016	PASS
95	95	100	100	100.0	44971	PASS
96	95	5	9	6.9	3116	PASS
173	174	0.00	2	0.9	283	PASS
174	95	50	100	71.2	32021	PASS
175	174	5	9	7.9	2541	PASS
176	174	95	101	95.8	30669	PASS
177	176	5	9	6.7	2055	PASS