

Surrogate Summary

SDG No.: **Q1283**

Client: **Tetra Tech NUS, Inc.**

Analytical Method: **SW8260-Low**

Lab Sample ID	Client ID	Parameter	Spike	Result	RecoveryQual	Limits	
						Low	High
Q1283-01	VPB192-HYD-20250131	1,2-Dichloroethane-d4	50	53.3	107	81	118
		Dibromofluoromethane	50	49.3	99	80	119
		Toluene-d8	50	49.5	99	89	112
		4-Bromofluorobenzene	50	51.0	102	85	114
Q1283-02	BP-VPB-192-TB-20250130	1,2-Dichloroethane-d4	50	53.5	107	81	118
		Dibromofluoromethane	50	49.4	99	80	119
		Toluene-d8	50	50.3	101	89	112
		4-Bromofluorobenzene	50	52.1	104	85	114
Q1283-03	BP-VPB-192-EB-20250131	1,2-Dichloroethane-d4	50	52.9	106	81	118
		Dibromofluoromethane	50	50.4	101	80	119
		Toluene-d8	50	50.0	100	89	112
		4-Bromofluorobenzene	50	51.8	104	85	114
Q1283-04	BP-VPB-192-GW-480-482	1,2-Dichloroethane-d4	50	54.4	109	81	118
		Dibromofluoromethane	50	50.3	101	80	119
		Toluene-d8	50	50.5	101	89	112
		4-Bromofluorobenzene	50	53.6	107	85	114
Q1283-05	BP-VPB-192-GW-500-502	1,2-Dichloroethane-d4	50	53.0	106	81	118
		Dibromofluoromethane	50	50.5	101	80	119
		Toluene-d8	50	49.6	99	89	112
		4-Bromofluorobenzene	50	51.6	103	85	114
VX0211WBL01	VX0211WBL01	1,2-Dichloroethane-d4	50	53.5	107	81	118
		Dibromofluoromethane	50	51.0	102	80	119
		Toluene-d8	50	50.3	100	89	112
		4-Bromofluorobenzene	50	52.4	105	85	114
VX0211WBS01	VX0211WBS01	1,2-Dichloroethane-d4	50	43.8	88	81	118
		Dibromofluoromethane	50	45.5	91	80	119
		Toluene-d8	50	46.2	92	89	112
		4-Bromofluorobenzene	50	47.4	95	85	114
VX0211WBSD0	VX0211WBSD01	1,2-Dichloroethane-d4	50	49.4	99	81	118
		Dibromofluoromethane	50	48.8	98	80	119
		Toluene-d8	50	49.0	98	89	112
		4-Bromofluorobenzene	50	49.6	99	85	114