

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: TETR06
 Lab Code: CHEM Case No.: Q1283 SAS No.: Q1283 SDG NO.: Q1283
 Lab File ID: VX044893.D Date Analyzed: 02/11/2025
 Instrument ID: MSVOA_X Time Analyzed: 10:09
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	125474	5.54	221806	6.75	200705	10.05
UPPER LIMIT	250948	6.044	443612	7.251	401410	10.549
LOWER LIMIT	62737	5.044	110903	6.251	100353	9.549
EPA SAMPLE NO.						
VPB192-HYD-20250131	91747	5.54	188311	6.76	169014	10.05
BP-VPB-192-TB-20250130	94281	5.54	193913	6.76	175364	10.05
BP-VPB-192-EB-20250131	94090	5.54	191976	6.76	173093	10.05
BP-VPB-192-GW-480-482	92053	5.55	188552	6.76	173671	10.05
BP-VPB-192-GW-500-502	94241	5.54	190687	6.76	172611	10.05
VX0211WBL01	98496	5.54	199075	6.76	183423	10.05
VX0211WBS01	122388	5.54	213544	6.76	190238	10.05
VX0211WBSD01	116406	5.54	211404	6.76	186335	10.05

IS1 = Pentafluorobenzene
 IS2 = 1,4-Difluorobenzene
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.