

Continuing Calibration Report for SequenceID : FC020525AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FC068192.D

Aliphatic C9-C12	8486200.000	60.000	3.110	6.377	141436.667	138488.174	-2.129
Aliphatic C12-C16	5547848.000	40.000	6.378	9.761	138696.200	137713.071	-0.714
Aliphatic C16-C21	8097363.000	60.000	9.762	13.114	134956.050	135019.289	0.047
Aliphatic C21-C28	10470276.000	80.000	13.115	16.766	130878.450	127207.526	-2.886
Aliphatic C28-C40	13162033.000	120.000	16.767	21.570	109683.608	103638.860	-5.833
Aliphatic EPH	45763720.000	360.000	3.110	21.570	127121.444	123700.655	-2.765