

Continuing Calibration Report for SequenceID : FC020525AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FC068204.D

Aliphatic C9-C12	8760040.000	60.000	3.110	6.377	146000.667	138488.174	-5.425
Aliphatic C12-C16	5699246.000	40.000	6.378	9.761	142481.150	137713.071	-3.462
Aliphatic C16-C21	8210418.000	60.000	9.762	13.114	136840.300	135019.289	-1.349
Aliphatic C21-C28	10240163.000	80.000	13.115	16.766	128002.038	127207.526	-0.625
Aliphatic C28-C40	12674624.000	120.000	16.767	21.570	105621.867	103638.860	-1.913
Aliphatic EPH	45584491.000	360.000	3.110	21.570	126623.586	123700.655	-2.363