

SOLID EPH_F2 LABORATORY CONTROL SPIKE/LABORATORY CONTROL SPIKE DUPLICATE RECOVERY

Lab Name: Chemtech **Client:** Sciacca General Contractors, LLC
Lab Code: CHEM **Cas No:** Q1290 **SAS No :** Q1290 **SDG No:** Q1290
Sample No : PB166564BSD **Datafile:** FC068195.D
Client ID : PB166564BSD

COMPOUND	SPIKE ADDED mg/kg	LCS/LCSD CONCENTRATION mg/kg	% REC	Qual	RPD QC LIMITS	QC Limit Of RPD
Aliphatic C28-C40	30.0	22.3	74		2.3 (40-140)	50
Aliphatic C9-C28	99.9	71.0	72		1 (40-140)	50