

TPH GC INITIAL CALIBRATION SUMMARY

Lab Name: Chemtech Contract: SCIA01
 ProjectID: 305 Central Ave, West Caldwell
 Lab Code: CHEM Case No.: Q1290 SAS No.: Q1290 SDG No.: Q1290

Calibration Sequence : FG011325		Test : TPH GC	
Concentration (PPM)	Area Count	Reference Factor	File ID
850	127030037	149447	FG015057.D
340	40811567	120034	FG015058.D
170	22836827	134334	FG015059.D
85	11459490	134818	FG015060.D
1700	199696868	117469	FG015061.D
AVG RF : 131220		% RSD : 9.856	AVG RT : 15.0514