

TPH GC CONTINUING CALIBRATION SUMMARY

50 PPM TRPH STD

Lab Name: Chemtech Contract: SCIA01
ProjectID: 305 Central Ave, West Caldwell
Lab Code: CHEM Case No.: Q1290 SAS No.: Q1290 SDG No.: Q1290
DataFile: FG015359.D Analyst Name: YP\AJ Analyst Date: 02-07-2025

Conc. (PPM)	Area Count	RF	Average RF	%D
850	93941977	110520	131220	15.775