

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: SCIA01
 Lab Code: CHEM Case No.: Q1290 SAS No.: Q1290 SDG NO.: Q1290
 Lab File ID: VY021130.D Date Analyzed: 02/07/2025
 Instrument ID: MSVOA_Y Time Analyzed: 09:31
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) Y

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	195996	7.72	297405	8.62	257416	11.43
UPPER LIMIT	391992	8.219	594810	9.122	514832	11.926
LOWER LIMIT	97998	7.219	148703	8.122	128708	10.926
EPA SAMPLE NO.						
VOC	11811 *	7.71	17126 *	8.62	12326 *	11.42
VY0207SBL01	222650	7.72	403911	8.62	364645	11.43
VY0207SBS01	185601	7.71	293384	8.62	248895	11.42
VY0207SBSD01	168157	7.71	264979	8.62	227375	11.42

IS1 = Pentafluorobenzene
 IS2 = 1,4-Difluorobenzene
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.