

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: SCIA01
 Lab Code: CHEM Case No.: Q1290 SAS No.: Q1290 SDG NO.: Q1290
 Lab File ID: VY021130.D Date Analyzed: 02/07/2025
 Instrument ID: MSVOA_Y Time Analyzed: 09:31
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) Y

	IS4 AREA #	RT #				
12 HOUR STD	124211	13.352				
UPPER LIMIT	248422	13.852				
LOWER LIMIT	62105.5	12.852				
EPA SAMPLE NO.						
VOC	2084 *	13.35				
VY0207SBL01	133881	13.35				
VY0207SBS01	123887	13.35				
VY0207SBSD01	111644	13.35				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.