

FORM 15-IN
INITIAL CALIBRATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
 Lab Code: ACE Case No.: 51960 MA No.: _____ SDG No.: E1PM9
 Instrument ID: P4 Start Date: 02/17/2025
 Analytical Method: ICP-AES Run Batch: LB134744
 Concentration Units: ug/L

Analyte	True	Found	%D	True	Found	%D	True	Found	%D
Arsenic	0	-0	0	10	12	20	1250	1100	-12
Barium	0	0.010	0	200	190	-5	2500	2500	0
Cadmium	0	-0	0	5	5.1	2	625	630	1
Chromium	0	-0	0	10	11	10	3875	4100	6
Lead	0	0.0	0	10	8.4	-16	6250	6300	1
Selenium	0	0.010	0	35	30	-14	1250	1200	-4
Silver	0	0.0	0	10	10	0	312.5	300	-4

FORM 15-IN
INITIAL CALIBRATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
Lab Code: ACE Case No.: 51960 MA No.: SDG No.: E1PM9
Instrument ID: P4 Start Date: 02/17/2025
Analytical Method: ICP-AES Run Batch: LB134744
Concentration Units: ug/L

Analyte	True	Found	%D	True	Found	%D	True	Found	%D
Arsenic	2500	2300	-8	5000	4800	-4	10000	10000	0
Barium	5000	5100	2	10000	10000	0	20000	20000	0
Cadmium	1250	1300	4	2500	2500	0	5000	5000	0
Chromium	7750	8100	5	15500	15000	-3	31000	30000	-3
Lead	12500	12000	-4	25000	25000	0	50000	50000	0
Selenium	2500	2400	-4	5000	4900	-2	10000	10000	0
Silver	625	610	-2	1250	1200	-4	2500	2600	4

FORM 15-IN
INITIAL CALIBRATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
Lab Code: ACE Case No.: 51960 MA No.: SDG No.: E1PM9
Instrument ID: P4 Start Date: 02/17/2025
Analytical Method: ICP-AES Run Batch: LB134744
Concentration Units: ug/L

Analyte	True	Found	%D	True	Found	%D	True	Found	%D
Arsenic	0	0.0	0						
Barium	0	0.0	0						
Cadmium	0	0.0	0						
Chromium	0	0.0	0						
Lead	0	0.0	0						
Selenium	0	0.0	0						
Silver	0	0.0	0						