

FORM 2 - IN

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011Lab Code: ACE Case No.: 51960 MA No. : SDG No.: E1PM9Initial Calibration Verification Source : MP84446Continuing Calibration Verification Source : MP84392Run Batch: LB134744 Analytical Method: ICP-AESConcentration Units: $\mu\text{g/L}$

	Initial Calibration Verification				Continuing Calibration Verification						
	ID: ICV001				ID: CCV001				ID: CCV002		
Analyte	True	Found	%R	%RSD	True	Found	%R	%RSD	Found	%R	%RSD
Arsenic	1000	1000	102	1	5000	4800	97	0	4700	94	1
Barium	520	510	98	0	10000	11000	106	0	10000	100	2
Cadmium	510	500	99	0	2500	2500	98	0	2500	99	0
Chromium	520	550	106	0	15500	16000	100	0	16000	100	0
Lead	1000	990	99	0	25000	24000	97	0	24000	97	0
Selenium	1000	920	92	1	5000	4900	98	0	4700	94	1
Silver	250	230	93	0	1250	1300	102	0	1300	101	0