

FORM 2 - IN

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011

Lab Code: ACE Case No.: 51960 MA No. : SDG No.: E1PM9

Initial Calibration Verification Source : MP84538

Continuing Calibration Verification Source : MP84540

Run Batch: LB134732 Analytical Method: CVAA

Concentration Units: µg/L

	Initial Calibration Verification				Continuing Calibration Verification						
	ID: ICV006				ID: CCV076				ID: CCV077		
Analyte	True	Found	%R	%RSD	True	Found	%R	%RSD	Found	%R	%RSD
Mercury	4.0	3.6	90		5.0	5.2	103		5.2	104	