

LB134732


SFAM01.1

INSTRUMENT ID: CV1

Linear

μ Abs.:

41317



A= 0.0000e+000

B= 2.3879e-004 *slope*

C= 1.1238e-001 *intercept*

Rho= 0.9998800

Accept= Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	0.038	0.038	-313	0.000	-313					=
0.05	0.050										14
0.20	0.200	0.228	0.028	485	0.0 %	485					-3
2.50	2.500	2.424	-0.076	9681	0.0 %	9681					-1
5.00	5.000	4.943	-0.057	20229	0.0 %	20229					1
7.50	7.500	7.589	0.089	31310	0.0 %	31310					0
10.0	10.000	9.978	-0.022	41317	0.0 %	41317					

LB134732

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S00	-313	-		0 SFAM01.1	PPB	2/17/2025 11:44	Std
	0.2 S01	485	-		0.2 SFAM01.1	PPB	2/17/2025 11:46	Std
	2.5 S02	9681	-		2.5 SFAM01.1	PPB	2/17/2025 11:48	Std
	5 S03	20229	-		5 SFAM01.1	PPB	2/17/2025 11:50	Std
	7.5 S04	31310	-		7.5 SFAM01.1	PPB	2/17/2025 11:53	Std
	10 S05	41317	-		10 SFAM01.1	PPB	2/17/2025 12:01	Std
ICV006	ICV006	14687	3.6195 -		SFAM01.1	PPB	2/17/2025 12:15	SMPL
ICB006	ICB006	-180	0.0694 -		SFAM01.1	PPB	2/17/2025 12:17	SMPL
CCV076	CCV076	21113	5.1539 -		SFAM01.1	PPB	2/17/2025 12:20	SMPL
CCB076	CCB076	-239	0.0553 -		SFAM01.1	PPB	2/17/2025 12:22	SMPL
Q1292-01	E1PM9	-30	0.1052 -		SFAM01.1	PPB	2/17/2025 12:31	SMPL
Q1292-02	E1PM9D	11	0.115 -		SFAM01.1	PPB	2/17/2025 12:34	SMPL
Q1292-03	E1PM9S	4581	1.2063 -		SFAM01.1	PPB	2/17/2025 12:36	SMPL
PB166746BL	PBW746	-57	0.0988 -		SFAM01.1	PPB	2/17/2025 12:39	SMPL
PB166660TB	LEB660	-15	0.1088 -		SFAM01.1	PPB	2/17/2025 12:44	SMPL
CCV077	CCV077	21303	5.1993 -		SFAM01.1	PPB	2/17/2025 12:47	SMPL
CCB077	CCB077	-95	0.0897 -		SFAM01.1	PPB	2/17/2025 12:49	SMPL