

Analysis Method: 9045D
Parameter: pH
Run Number: LB134597
BalanceID: WC SC-7

Analyst By : jignesh
Supervisor Review By : Iwona
Slope : 98.3
pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3107
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER, COLOR CD 475ML	W3094
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3161
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	02/06/2025	08:30
2	CAL2	1	Water	NA	NA	20.2	7.00	02/06/2025	08:31
3	CAL3	1	Water	NA	NA	20.3	10.02	02/06/2025	08:35
4	ICV	1	Water	NA	NA	20.3	7.00	02/06/2025	08:40
5	CCV1	1	Water	NA	NA	20.2	2.01	02/06/2025	08:45
6	Q1295-01	1	Solid	20.02	20	20.9	6.74	02/06/2025	08:55
7	Q1295-01DUP	1	Solid	20.03	20	20.8	6.75	02/06/2025	08:56
8	CCV2	1	Water	NA	NA	20.3	12.02	02/06/2025	09:00

17124597

WORKLIST(Hardcopy Internal Chain)

WorkList Name : ph q1295 WorkList ID : 187511 Department : Wet-Chemistry Date : 02-06-2025 08:21:45

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q1295-01	TAPFTA-SB01D-020425-00-T1	Solid	pH	Cool 4 deg C	WEST04	N41	02/04/2025	9045D

Date/Time 02/06/25 08:25
 Raw Sample Received by: JH (WOC)
 Raw Sample Relinquished by: CSM

Date/Time 02/06/25 10:30
 Raw Sample Received by: CSM
 Raw Sample Relinquished by: JH (WOC)