

Initial and Continuing Calibration Blank Summary

Client: Weston Solutions

SDG No.: Q1295

Project: Ft Meade Tipton Airfield Parcel RI - PO 0111169

RunNo.: LB134598

Analyte	Units	Result	Acceptance Limits	Conc Qual	MDL	RDL	Analysis Date
Sample ID: ICB1							
Bromide	mg/L	< 1.0000	1.0000	U	0.034	2	01/20/2025
Chloride	mg/L	< 0.3000	0.3000	U	0.011	0.6	01/20/2025
Fluoride	mg/L	< 0.2000	0.2000	U	0.057	0.4	01/20/2025
Nitrite	mg/L	< 0.3000	0.3000	U	0.011	0.6	01/20/2025
Nitrate	mg/L	< 0.2500	0.2500	U	0.0034	0.5	01/20/2025
Sulfate	mg/L	< 1.5000	1.5000	U	0.032	3	01/20/2025
Orthophosphate as P	mg/L	< 0.5000	0.5000	U	0.079	1	01/20/2025
Sample ID: CCB1							
Bromide	mg/L	< 1.0000	1.0000	U	0.034	2	02/05/2025
Chloride	mg/L	< 0.3000	0.3000	U	0.011	0.6	02/05/2025
Fluoride	mg/L	< 0.2000	0.2000	U	0.057	0.4	02/05/2025
Nitrite	mg/L	< 0.3000	0.3000	U	0.011	0.6	02/05/2025
Nitrate	mg/L	< 0.2500	0.2500	U	0.0034	0.5	02/05/2025
Sulfate	mg/L	< 1.5000	1.5000	U	0.032	3	02/05/2025
Orthophosphate as P	mg/L	< 0.5000	0.5000	U	0.079	1	02/05/2025
Sample ID: CCB2							
Bromide	mg/L	< 1.0000	1.0000	U	0.034	2	02/05/2025
Chloride	mg/L	< 0.3000	0.3000	U	0.011	0.6	02/05/2025
Fluoride	mg/L	< 0.2000	0.2000	U	0.057	0.4	02/05/2025
Nitrite	mg/L	< 0.3000	0.3000	U	0.011	0.6	02/05/2025
Nitrate	mg/L	< 0.2500	0.2500	U	0.0034	0.5	02/05/2025
Sulfate	mg/L	< 1.5000	1.5000	U	0.032	3	02/05/2025
Orthophosphate as P	mg/L	< 0.5000	0.5000	U	0.079	1	02/05/2025

Initial and Continuing Calibration Blank Summary

Client: Weston Solutions	SDG No.: Q1295
Project: Ft Meade Tipton Airfield Parcel RI - PO 0111169	RunNo.: LB134653

Analyte	Units	Result	Acceptance Limits	Conc Qual	MDL	RDL	Analysis Date
Sample ID: CCB1 TOC	mg/L	< 125.0000	125.0000	U	22.3	250	02/10/2025
Sample ID: CCB2 TOC	mg/L	< 125.0000	125.0000	U	22.3	250	02/10/2025
Sample ID: ICB1 TOC	mg/L	< 125.0000	125.0000	U	22.3	250	11/12/2024