

Laboratory Control Sample Summary

Client:	Weston Solutions	SDG No.:	Q1295
Project:	Ft Meade Tipton Airfield Parcel RI - PO 0111169	Run No.:	LB134598

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB134598BSS							
Bromide	mg/Kg	200	206		103	1	90-110	02/05/2025
Chloride	mg/Kg	60	61.6		103	1	90-110	02/05/2025
Fluoride	mg/Kg	40	40.8		102	1	90-110	02/05/2025
Nitrite	mg/Kg	60	61.4		102	1	90-110	02/05/2025
Nitrate	mg/Kg	50	51.4		103	1	90-110	02/05/2025
Sulfate	mg/Kg	300	305		102	1	90-110	02/05/2025
Orthophosphate as P	mg/Kg	100	101		101	1	90-110	02/05/2025

Laboratory Control Sample Summary

Client: Weston Solutions

SDG No.: Q1295

Project: Ft Meade Tipton Airfield Parcel RI - PO 0111169

Run No.: LB134653

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB134653BSS							
TOC	mg/Kg	1000	1010		101	1	90-110	02/10/2025