



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client:	<u>Weston Solutions</u>				SDG No.:	<u>Q1295</u>				
Contract:	<u>WEST04</u>	Lab Code:	<u>CHEM</u>	Case No.:	<u>Q1295</u>	SAS No.:	<u>Q1295</u>			
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	LOD	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB89	Mercury	0.20	+/-0.20	U	0.16	0.20	CV	02/06/2025	13:42	LB134603

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Weston Solutions

SDG No.: Q1295

Contract: WEST04

Lab Code: CHEM

Case No.: Q1295

SAS No.: Q1295

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	LOD	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB33	Mercury	0.20	+/-0.20	U	0.16	0.20	CV	02/06/2025	13:47	LB134603
CCB34	Mercury	0.20	+/-0.20	U	0.16	0.20	CV	02/06/2025	14:22	LB134603
CCB35	Mercury	0.20	+/-0.20	U	0.16	0.20	CV	02/06/2025	14:54	LB134603
CCB36	Mercury	0.20	+/-0.20	U	0.16	0.20	CV	02/06/2025	15:17	LB134603

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: <u>Weston Solutions</u>		SDG No.: <u>Q1295</u>								
Contract: <u>WEST04</u>	Lab Code: <u>CHEM</u>	Case No.: <u>Q1295</u>	SAS No.: <u>Q1295</u>							
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	LOD	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB01	Aluminum	20.0	+/-20.0	U	10.0	20.0	P	02/17/2025	12:55	LB134737
	Antimony	2.00	+/-2.00	U	0.25	2.00	P	02/17/2025	12:55	LB134737
	Arsenic	1.00	+/-1.00	U	0.25	1.00	P	02/17/2025	12:55	LB134737
	Barium	10.0	+/-10.0	U	1.25	10.0	P	02/17/2025	12:55	LB134737
	Beryllium	1.00	+/-1.00	U	0.25	1.00	P	02/17/2025	12:55	LB134737
	Cadmium	1.00	+/-1.00	U	0.50	1.00	P	02/17/2025	12:55	LB134737
	Calcium	500	+/-500	U	190	500	P	02/17/2025	12:55	LB134737
	Chromium	2.00	+/-2.00	U	0.75	2.00	P	02/17/2025	12:55	LB134737
	Cobalt	1.00	+/-1.00	U	0.25	1.00	P	02/17/2025	12:55	LB134737
	Copper	2.00	+/-2.00	U	1.50	2.00	P	02/17/2025	12:55	LB134737
	Iron	50.0	+/-50.0	U	25.0	50.0	P	02/17/2025	12:55	LB134737
	Lead	1.00	+/-1.00	U	0.75	1.00	P	02/17/2025	12:55	LB134737
	Magnesium	500	+/-500	U	190	500	P	02/17/2025	12:55	LB134737
	Manganese	1.00	+/-1.00	U	0.75	1.00	P	02/17/2025	12:55	LB134737
	Nickel	1.00	+/-1.00	U	0.25	1.00	P	02/17/2025	12:55	LB134737
	Potassium	500	+/-500	U	190	500	P	02/17/2025	12:55	LB134737
	Selenium	5.00	+/-5.00	U	4.50	5.00	P	02/17/2025	12:55	LB134737
	Silver	1.00	+/-1.00	U	0.50	1.00	P	02/17/2025	12:55	LB134737
	Sodium	500	+/-500	U	190	500	P	02/17/2025	12:55	LB134737
	Thallium	1.00	+/-1.00	U	0.50	1.00	P	02/17/2025	12:55	LB134737
	Vanadium	5.00	+/-5.00	U	0.25	5.00	P	02/17/2025	12:55	LB134737
	Zinc	5.00	+/-5.00	U	1.50	5.00	P	02/17/2025	12:55	LB134737

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: <u>Weston Solutions</u>		SDG No.: <u>Q1295</u>								
Contract: <u>WEST04</u>	Lab Code: <u>CHEM</u>	Case No.: <u>Q1295</u>	SAS No.: <u>Q1295</u>							
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	LOD	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB01	Aluminum	20.0	+/-20.0	U	10.0	20.0	P	02/17/2025	13:16	LB134737
	Antimony	0.13	+/-2.00	J	0.25	2.00	P	02/17/2025	13:16	LB134737
	Arsenic	1.00	+/-1.00	U	0.25	1.00	P	02/17/2025	13:16	LB134737
	Barium	0.39	+/-10.0	J	1.25	10.0	P	02/17/2025	13:16	LB134737
	Beryllium	0.27	+/-1.00	J	0.25	1.00	P	02/17/2025	13:16	LB134737
	Cadmium	1.00	+/-1.00	U	0.50	1.00	P	02/17/2025	13:16	LB134737
	Calcium	500	+/-500	U	190	500	P	02/17/2025	13:16	LB134737
	Chromium	2.00	+/-2.00	U	0.75	2.00	P	02/17/2025	13:16	LB134737
	Cobalt	1.00	+/-1.00	U	0.25	1.00	P	02/17/2025	13:16	LB134737
	Copper	2.00	+/-2.00	U	1.50	2.00	P	02/17/2025	13:16	LB134737
	Iron	50.0	+/-50.0	U	25.0	50.0	P	02/17/2025	13:16	LB134737
	Lead	0.20	+/-1.00	J	0.75	1.00	P	02/17/2025	13:16	LB134737
	Magnesium	500	+/-500	U	190	500	P	02/17/2025	13:16	LB134737
	Manganese	1.00	+/-1.00	U	0.75	1.00	P	02/17/2025	13:16	LB134737
	Nickel	1.00	+/-1.00	U	0.25	1.00	P	02/17/2025	13:16	LB134737
	Potassium	500	+/-500	U	190	500	P	02/17/2025	13:16	LB134737
	Selenium	5.00	+/-5.00	U	4.50	5.00	P	02/17/2025	13:16	LB134737
	Silver	1.00	+/-1.00	U	0.50	1.00	P	02/17/2025	13:16	LB134737
	Sodium	500	+/-500	U	190	500	P	02/17/2025	13:16	LB134737
	Thallium	1.00	+/-1.00	U	0.50	1.00	P	02/17/2025	13:16	LB134737
	Vanadium	5.00	+/-5.00	U	0.25	5.00	P	02/17/2025	13:16	LB134737
	Zinc	5.00	+/-5.00	U	1.50	5.00	P	02/17/2025	13:16	LB134737
CCB02	Aluminum	20.0	+/-20.0	U	10.0	20.0	P	02/17/2025	14:04	LB134737
	Antimony	2.00	+/-2.00	U	0.25	2.00	P	02/17/2025	14:04	LB134737
	Arsenic	1.00	+/-1.00	U	0.25	1.00	P	02/17/2025	14:04	LB134737
	Barium	10.0	+/-10.0	U	1.25	10.0	P	02/17/2025	14:04	LB134737
	Beryllium	1.00	+/-1.00	U	0.25	1.00	P	02/17/2025	14:04	LB134737
	Cadmium	1.00	+/-1.00	U	0.50	1.00	P	02/17/2025	14:04	LB134737
	Calcium	500	+/-500	U	190	500	P	02/17/2025	14:04	LB134737
	Chromium	2.00	+/-2.00	U	0.75	2.00	P	02/17/2025	14:04	LB134737
	Cobalt	1.00	+/-1.00	U	0.25	1.00	P	02/17/2025	14:04	LB134737
	Copper	2.00	+/-2.00	U	1.50	2.00	P	02/17/2025	14:04	LB134737
	Iron	50.0	+/-50.0	U	25.0	50.0	P	02/17/2025	14:04	LB134737
	Lead	1.00	+/-1.00	U	0.75	1.00	P	02/17/2025	14:04	LB134737
	Magnesium	500	+/-500	U	190	500	P	02/17/2025	14:04	LB134737
	Manganese	1.00	+/-1.00	U	0.75	1.00	P	02/17/2025	14:04	LB134737
	Nickel	1.00	+/-1.00	U	0.25	1.00	P	02/17/2025	14:04	LB134737
	Potassium	500	+/-500	U	190	500	P	02/17/2025	14:04	LB134737
	Selenium	5.00	+/-5.00	U	4.50	5.00	P	02/17/2025	14:04	LB134737

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: <u>Weston Solutions</u>		SDG No.: <u>Q1295</u>								
Contract: <u>WEST04</u>	Lab Code: <u>CHEM</u>	Case No.: <u>Q1295</u>	SAS No.: <u>Q1295</u>							
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	LOD	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB02	Silver	1.00	+/-1.00	U	0.50	1.00	P	02/17/2025	14:04	LB134737
	Sodium	500	+/-500	U	190	500	P	02/17/2025	14:04	LB134737
	Thallium	1.00	+/-1.00	U	0.50	1.00	P	02/17/2025	14:04	LB134737
	Vanadium	5.00	+/-5.00	U	0.25	5.00	P	02/17/2025	14:04	LB134737
	Zinc	5.00	+/-5.00	U	1.50	5.00	P	02/17/2025	14:04	LB134737
CCB03	Aluminum	20.0	+/-20.0	U	10.0	20.0	P	02/17/2025	14:34	LB134737
	Antimony	2.00	+/-2.00	U	0.25	2.00	P	02/17/2025	14:34	LB134737
	Arsenic	1.00	+/-1.00	U	0.25	1.00	P	02/17/2025	14:34	LB134737
	Barium	10.0	+/-10.0	U	1.25	10.0	P	02/17/2025	14:34	LB134737
	Beryllium	1.00	+/-1.00	U	0.25	1.00	P	02/17/2025	14:34	LB134737
	Cadmium	1.00	+/-1.00	U	0.50	1.00	P	02/17/2025	14:34	LB134737
	Calcium	500	+/-500	U	190	500	P	02/17/2025	14:34	LB134737
	Chromium	2.00	+/-2.00	U	0.75	2.00	P	02/17/2025	14:34	LB134737
	Cobalt	1.00	+/-1.00	U	0.25	1.00	P	02/17/2025	14:34	LB134737
	Copper	2.00	+/-2.00	U	1.50	2.00	P	02/17/2025	14:34	LB134737
	Iron	50.0	+/-50.0	U	25.0	50.0	P	02/17/2025	14:34	LB134737
	Lead	1.00	+/-1.00	U	0.75	1.00	P	02/17/2025	14:34	LB134737
	Magnesium	500	+/-500	U	190	500	P	02/17/2025	14:34	LB134737
	Manganese	1.00	+/-1.00	U	0.75	1.00	P	02/17/2025	14:34	LB134737
	Nickel	1.00	+/-1.00	U	0.25	1.00	P	02/17/2025	14:34	LB134737
	Potassium	500	+/-500	U	190	500	P	02/17/2025	14:34	LB134737
	Selenium	5.00	+/-5.00	U	4.50	5.00	P	02/17/2025	14:34	LB134737
	Silver	1.00	+/-1.00	U	0.50	1.00	P	02/17/2025	14:34	LB134737
	Sodium	500	+/-500	U	190	500	P	02/17/2025	14:34	LB134737
	Thallium	1.00	+/-1.00	U	0.50	1.00	P	02/17/2025	14:34	LB134737
	Vanadium	5.00	+/-5.00	U	0.25	5.00	P	02/17/2025	14:34	LB134737
	Zinc	5.00	+/-5.00	U	1.50	5.00	P	02/17/2025	14:34	LB134737