

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: Weston Solutions

SDG No.: Q1295

Contract: WEST04

Lab Code: CHEM

Case No.: Q1295

SAS No.: Q1295

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB166584BS Mercury	mg/Kg	0.26	0.26		101	80 - 124	CV

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: Weston Solutions **SDG No.:** Q1295
Contract: WEST04 **Lab Code:** CHEM **Case No.:** Q1295 **SAS No.:** Q1295

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB166734BS							
Aluminum	mg/Kg	430	441		103	78 - 124	P
Antimony	mg/Kg	21.7	22.7		105	72 - 124	P
Arsenic	mg/Kg	21.7	22.6		104	82 - 118	P
Barium	mg/Kg	110	115		104	86 - 116	P
Beryllium	mg/Kg	21.7	22.5		104	80 - 120	P
Cadmium	mg/Kg	21.7	23.3		107	84 - 116	P
Calcium	mg/Kg	2200	2320		106	86 - 118	P
Chromium	mg/Kg	21.7	22.7		105	83 - 119	P
Cobalt	mg/Kg	21.7	22.5		104	84 - 115	P
Copper	mg/Kg	220	228		104	84 - 119	P
Iron	mg/Kg	1100	1200		109	81 - 124	P
Lead	mg/Kg	110	111		101	84 - 118	P
Magnesium	mg/Kg	2200	2390		109	80 - 123	P
Manganese	mg/Kg	220	228		104	85 - 116	P
Nickel	mg/Kg	21.7	21.4		99	84 - 119	P
Potassium	mg/Kg	1100	1160		106	85 - 119	P
Selenium	mg/Kg	21.7	22.8		105	80 - 119	P
Silver	mg/Kg	21.7	23.5		108	83 - 118	P
Sodium	mg/Kg	2200	2350		107	79 - 125	P
Thallium	mg/Kg	21.7	22.4		103	83 - 118	P
Vanadium	mg/Kg	21.7	21.5		99	82 - 116	P
Zinc	mg/Kg	220	231		105	82 - 119	P