

FORM 8B

ICP-MS INTERNAL STANDARD RELATIVE INTENSITY SUMMARY

Lab Name: Weston Solutions

Contract: WEST04

Lab Code: CHEM Case no.: Q1295

Sas No.: Q1295 SDG No.: Q1295

Instrument ID: P8

Start Date : 02/17/2025

Run Number: LB134737

End Date : 02/17/2025

Lab SampleID	Client SampleID	Time	Internal Standard %RI For: Non-Collision Cell									
			Element 165Ho	Q	Element 209Bi	Q	Element	Q	Element	Q	Element	Q
S0	S0	1215	100		100							
S2	S2	1219	101		101							
S3	S3	1222	102		100							
S4	S4	1225	102		99							
S5	S5	1228	101		98							
S6	S6	1231	101		96							
S7	S7	1234	101		93							
S8	S8	1237	93		82							
ICV01	ICV01	1248	104		102							
LLICV	LLICV	1252	103		102							
ICB01	ICB01	1255	101		101							
ICSA01	ICSA01	1258	101		92							
ICSAB01	ICSAB01	1302	101		93							
CCV01	CCV01	1313	97		87							
CCB01	CCB01	1316	97		96							
CRI	CRI	1319	100		99							
PB166734BL	PB166734BL	1323	99		97							
PB166734BS	PB166734BS	1326	98		93							
Q1295-01	TAPFTA-SB01I	1329	98		97							
Q1295-01DUP	TAPFTA-SB01I	1332	98		95							
Q1295-01L	TAPFTA-SB01I	1336	99		98							
Q1295-01MS	TAPFTA-SB01I	1339	96		91							
Q1295-01MSD	TAPFTA-SB01I	1342	98		93							
Q1295-01A	TAPFTA-SB01I	1345	98		95							
CCV02	CCV02	1357	93		82							
CCB02	CCB02	1404	98		95							
Q1295-01	TAPFTA-SB01I	1410	100		99							
Q1295-01DUP	TAPFTA-SB01I	1414	102		98							
Q1295-01L	TAPFTA-SB01I	1417	101		98							
Q1295-01MS	TAPFTA-SB01I	1420	100		97							
Q1295-01MSD	TAPFTA-SB01I	1423	100		97							
Q1295-01A	TAPFTA-SB01I	1426	101		98							
CCV03	CCV03	1430	96		86							
CCB03	CCB03	1434	101		99							

Internal Standard %RI Limit: 30 -120

FORM 8B

ICP-MS INTERNAL STANDARD RELATIVE INTENSITY SUMMARY

Lab Name: Weston Solutions
Lab Code: CHEM Case no.: Q1295
Instrument ID: P8
Run Number: LB134737

Contract: WEST04
Sas No.: Q1295 SDG No.: Q1295
Start Date : 02/17/2025
End Date : 02/17/2025

Lab SampleID	Client SampleID	Time	Internal Standard %RI For: Collision Cell									
			Element 209Bi	Q	Element	Q	Element	Q	Element	Q	Element	Q
S0	S0	1215	100									
S2	S2	1219	99									
S3	S3	1222	99									
S4	S4	1225	99									
S5	S5	1228	97									
S6	S6	1231	96									
S7	S7	1234	94									
S8	S8	1237	83									
ICV01	ICV01	1248	102									
LLICV	LLICV	1252	102									
ICB01	ICB01	1255	101									
ICSA01	ICSA01	1258	92									
ICSAB01	ICSAB01	1302	93									
CCV01	CCV01	1313	89									
CCB01	CCB01	1316	101									
CRI	CRI	1319	100									
PB166734BL	PB166734BL	1323	99									
PB166734BS	PB166734BS	1326	93									
Q1295-01	TAPFTA-SB01I	1329	98									
Q1295-01DUP	TAPFTA-SB01I	1332	98									
Q1295-01L	TAPFTA-SB01I	1336	98									
Q1295-01MS	TAPFTA-SB01I	1339	96									
Q1295-01MSD	TAPFTA-SB01I	1342	96									
Q1295-01A	TAPFTA-SB01I	1345	96									
CCV02	CCV02	1357	86									
CCB02	CCB02	1404	98									
Q1295-01	TAPFTA-SB01I	1410	101									
Q1295-01DUP	TAPFTA-SB01I	1414	103									
Q1295-01L	TAPFTA-SB01I	1417	101									
Q1295-01MS	TAPFTA-SB01I	1420	99									
Q1295-01MSD	TAPFTA-SB01I	1423	99									
Q1295-01A	TAPFTA-SB01I	1426	99									
CCV03	CCV03	1430	89									
CCB03	CCB03	1434	102									

Internal Standard %RI Limit: 30 -120