

Metals
- 4 -
INTERFERENCE CHECK SAMPLE

Client: Aramark Uniforms **SDG No.:** Q1300
Contract: ARAM01 **Lab Code:** CHEM **Case No.:** Q1300 **SAS No.:** Q1300
ICS Source: EPA **Instrument ID:** P4

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSA01	Cadmium	2.97	1.0	297	-5	7	02/07/2025	12:40	LB134636
	Copper	9.83	2.0	492	-18	22	02/07/2025	12:40	LB134636
	Lead	8.02			-12	12	02/07/2025	12:40	LB134636
	Nickel	3.82	2.0	191	-38	42	02/07/2025	12:40	LB134636
	Zinc	5.22			-40	40	02/07/2025	12:40	LB134636
ICSAB01	Cadmium	1020	972	105	826	1120	02/07/2025	12:44	LB134636
	Copper	519	511	102	434	588	02/07/2025	12:44	LB134636
	Lead	56.3	49.0	115	37	61	02/07/2025	12:44	LB134636
	Nickel	1020	954	107	810	1100	02/07/2025	12:44	LB134636
	Zinc	1070	952	112	809	1095	02/07/2025	12:44	LB134636
ICSA	Cadmium	4.47	1.0	447	-5	7	02/07/2025	12:48	LB134636
	Copper	-15.3	2.0	765	-18	22	02/07/2025	12:48	LB134636
	Lead	-20.9			-12	12	02/07/2025	12:48	LB134636
	Nickel	4.48	2.0	224	-38	42	02/07/2025	12:48	LB134636
	Zinc	-0.68			-40	40	02/07/2025	12:48	LB134636
ICSAB	Cadmium	1040	972	107	826	1120	02/07/2025	12:52	LB134636
	Copper	526	511	103	434	588	02/07/2025	12:52	LB134636
	Lead	30.4	49.0	62	37	61	02/07/2025	12:52	LB134636
	Nickel	1040	954	109	810	1100	02/07/2025	12:52	LB134636
	Zinc	1060	952	111	809	1095	02/07/2025	12:52	LB134636