

Laboratory Control Sample Summary

Client: Parkway Generation Operating LLC

SDG No.: Q1301

Project: Linden Generating Station - 402

Run No.: LB134583

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB134583BS							
TPH	mg/L	20.0	16.9		84	1	78-114	02/05/2025

Laboratory Control Sample Summary

Client: Parkway Generation Operating LLC

SDG No.: Q1301

Project: Linden Generating Station - 402

Run No.: LB134583

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB134583BSD							
TPH	mg/L	20.0	17.0		85	1	78-114	02/05/2025