

Hit Summary Sheet
SW-846

SDG No.: Q1302
Client: GFE LLC

Sample ID	Client ID	Matrix	Parameter	Concentration	C	MDL	RDL	Units
Client ID:	SV1							
Q1302-01	SV1	Air	1,1,1-Trichloroethane	8.18		0.22	0.65	ug/m3
Q1302-01	SV1	Air	Trichloroethene	0.86		0.38	0.64	ug/m3
Q1302-01	SV1	Air	Tetrachloroethene	14.2		0.41	0.81	ug/m3
			Total Voc :	23.3				
			Total Concentration:	23.3				
Client ID:	SV2							
Q1302-02	SV2	Air	1,1,1-Trichloroethane	4.26		0.22	0.65	ug/m3
Q1302-02	SV2	Air	Tetrachloroethene	8.14		0.41	0.81	ug/m3
			Total Voc :	12.4				
			Total Concentration:	12.4				



SAMPLE DATA

A

B

C

D

Report of Analysis

Client:	GFE LLC	Date Collected:	02/05/25
Project:	55 Contant Ave Lodi, NJ	Date Received:	02/05/25
Client Sample ID:	SV1	SDG No.:	Q1302
Lab Sample ID:	Q1302-01	Matrix:	Air
Analytical Method:	TO-15	Test:	VOCMS Group3
Sample Wt/Vol:	400 Units: mL		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VL041982.D	4		02/10/25 12:04	VL021025

CAS Number	Parameter	Conc. ppbv	Conc. ug/M3	Qualifier	MDL	LOQ / CRQL	Units
TARGETS							
75-01-4	Vinyl Chloride	0.060	0.15	U	0.15	0.31	ug/m3
75-35-4	1,1-Dichloroethene	0.56	2.22	U	2.22	7.93	ug/m3
156-59-2	cis-1,2-Dichloroethene	0.36	1.43	U	1.43	7.93	ug/m3
71-55-6	1,1,1-Trichloroethane	1.50	8.18		0.22	0.65	ug/m3
79-01-6	Trichloroethene	0.16	0.86		0.38	0.64	ug/m3
127-18-4	Tetrachloroethene	2.10	14.2		0.41	0.81	ug/m3
SURROGATES							
460-00-4	1-Bromo-4-Fluorobenzene	9.80			65 - 135	98%	SPK: 10
INTERNAL STANDARDS							
74-97-5	Bromochloromethane	156000			2.781		
540-36-3	1,4-Difluorobenzene	428000			3.949		
3114-55-4	Chlorobenzene-d5	390000			8.872		

U = Not Detected

RL = Reporting Limit

MDL = Method Detection Limit

E = Value Exceeds Calibration Range

D = Dilution

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

Q = indicates LCS control criteria did not meet requirements

Report of Analysis

Client:	GFE LLC	Date Collected:	02/05/25
Project:	55 Contant Ave Lodi, NJ	Date Received:	02/05/25
Client Sample ID:	SV2	SDG No.:	Q1302
Lab Sample ID:	Q1302-02	Matrix:	Air
Analytical Method:	TO-15	Test:	VOCMS Group3
Sample Wt/Vol:	400 Units: mL		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VL041983.D	4		02/10/25 12:35	VL021025

CAS Number	Parameter	Conc. ppbv	Conc. ug/M3	Qualifier	MDL	LOQ / CRQL	Units
TARGETS							
75-01-4	Vinyl Chloride	0.060	0.15	U	0.15	0.31	ug/m3
75-35-4	1,1-Dichloroethene	0.56	2.22	U	2.22	7.93	ug/m3
156-59-2	cis-1,2-Dichloroethene	0.36	1.43	U	1.43	7.93	ug/m3
71-55-6	1,1,1-Trichloroethane	0.78	4.26		0.22	0.65	ug/m3
79-01-6	Trichloroethene	0.070	0.38	U	0.38	0.64	ug/m3
127-18-4	Tetrachloroethene	1.20	8.14		0.41	0.81	ug/m3
SURROGATES							
460-00-4	1-Bromo-4-Fluorobenzene	9.40			65 - 135	94%	SPK: 10
INTERNAL STANDARDS							
74-97-5	Bromochloromethane	157000			2.784		
540-36-3	1,4-Difluorobenzene	438000			3.956		
3114-55-4	Chlorobenzene-d5	397000			8.875		

U = Not Detected

RL = Reporting Limit

MDL = Method Detection Limit

E = Value Exceeds Calibration Range

D = Dilution

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

Q = indicates LCS control criteria did not meet requirements

LAB CHRONICLE

OrderID:	Q1302	OrderDate:	2/5/2025 1:24:00 PM
Client:	GFE LLC	Project:	55 Contant Ave Lodi, NJ
Contact:	Frank Galdun	Location:	Air Lab,VOA Lab

LabID	ClientID	Matrix	Test	Method	Sample Date	Prep Date	Anal Date	Received
Q1302-01	SV1	Air	VOCMS Group3	TO-15	02/05/25		02/10/25	02/05/25
Q1302-02	SV2	Air	VOCMS Group3	TO-15	02/05/25		02/10/25	02/05/25