

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: GFEL01

Lab Code: CHEM Case No.: Q1302 SAS No.: Q1302 SDG No.: Q1302

Instrument ID: MSVOA_L Calibration Date/Time: 02/10/2025 08:32

Lab File ID: VL041977.D Init. Calib. Date(s): 01/14/2025 01/14/2025

Heated Purge: (Y/N) N Init. Calib. Time(s): 12:55 16:05

GC Column: RTX-1 ID: 0.32 (mm)

COMPOUND	RRF	RRF010	MIN RRF	%D	MAX%D
Vinyl Chloride	0.519	0.540		4.05	30
1,1-Dichloroethene	0.501	0.493		-1.6	30
cis-1,2-Dichloroethene	1.349	1.330		-1.41	30
1,1,1-Trichloroethane	1.938	2.010		3.71	30
Trichloroethene	0.364	0.411		12.91	30
Tetrachloroethene	0.318	0.341		7.23	30
1-Bromo-4-Fluorobenzene	0.788	0.783		-0.63	30

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.