

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q1303

Client: GFE LLC

Analytical Method: SWTO-15

Datafile : VL041979.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VL0210ABS01	Vinyl Chloride	10	10.7	ppbv	107			70	130	
	Heptane	10	11.5	ppbv	115			70	130	
	1,1-Dichloroethene	10	9.50	ppbv	95			70	130	
	Cyclohexane	10	10.6	ppbv	106			70	130	
	cis-1,2-Dichloroethene	10	10.1	ppbv	101			70	130	
	1,1,1-Trichloroethane	10	10.4	ppbv	104			70	130	
	2,2,4-Trimethylpentane	10	12.2	ppbv	122			70	130	
	Benzene	10	11.7	ppbv	117			70	130	
	Trichloroethene	10	11.4	ppbv	114			70	130	
	Toluene	10	11.6	ppbv	116			70	130	
	1,2-Dibromoethane	10	11.8	ppbv	118			70	130	
	Tetrachloroethene	10	10.7	ppbv	107			70	130	
	Ethyl Benzene	10	11.8	ppbv	118			70	130	
	m/p-Xylene	20	24.3	ppbv	121			70	130	
	o-Xylene	10	12.2	ppbv	122			70	130	
	1,3,5-Trimethylbenzene	10	11.9	ppbv	119			70	130	
	1,2,4-Trimethylbenzene	10	12.2	ppbv	122			70	130	
	Naphthalene	10	10.6	ppbv	106			70	130	
	Hexane	10	11.2	ppbv	112			70	130	