

DIESEL RANGE ORGANICS INITIAL CALIBRATION SUMMARY

Lab Name: Chemtech Contract: ENTA05
 ProjectID: North Point - 4101 Arthur Kill Rd - E9306
 Lab Code: CHEM Case No.: Q1308 SAS No.: Q1308 SDG No.: Q1308

Calibration Sequence : FG011325		Test : Diesel Range Organics	
Concentration (PPM)	Area Count	Reference Factor	File ID
500	78963406	157927	FG015057.D
200	26178445	130892	FG015058.D
100	14314421	143144	FG015059.D
50	7127024	142540	FG015060.D
1000	126807043	126807	FG015061.D
AVG RF : 140262		% RSD : 8.691	AVG RT : 15.0514