

284 Sheffield Street, Mountainside, New Jersey 07092, Phone: 908 789 8900,

Date Received:

Test:

02/05/25

Diesel Range Organics

Fax: 908 789 8922

Report of Analysis

Client: ENTACT

NTACT Date Collected: 02/05/25

Project: North Point - 4101 Arthur Kill Rd - E9306

Client Sample ID: EX-11-TPH-9 SDG No.: Q1308

Lab Sample ID: Q1308-08 Matrix: SOIL

uL

Analytical Method: 8015D DRO % Solid: 90.3 Decanted:

Sample Wt/Vol: 30.05 Units: g Final Vol: 1 mL

Extraction Type: Injection Volume :

GPC Factor: PH:

Prep Method: SW3541

Soil Aliquot Vol:

File ID/Qc Batch: Dilution: Prep Date Date Analyzed Prep Batch ID

FG015350.D 200 02/06/25 09:15 02/07/25 9:43 PB166587

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS DRO	DRO	2540000		40900	36900	0 ug/kg
SURROGATES 16416-32-3	Tetracosane-d50	0.00	*	37 - 130	0%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit