

Continuing Calibration Report for SequenceID : FC020725AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FC068228.D

Aliphatic C9-C12	9027403.000	60.000	3.111	6.377	150456.717	138488.174	-8.642
Aliphatic C12-C16	5773976.000	40.000	6.378	9.762	144349.400	137713.071	-4.819
Aliphatic C16-C21	8274019.000	60.000	9.763	13.116	137900.317	135019.289	-2.134
Aliphatic C21-C28	10419562.000	80.000	13.117	16.769	130244.525	127207.526	-2.387
Aliphatic C28-C40	12969883.000	120.000	16.770	21.574	108082.358	103638.860	-4.287
Aliphatic EPH	46464843.000	360.000	3.111	21.574	129069.008	123700.655	-4.340