

Continuing Calibration Report for SequenceID : FC020725AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FC068240.D

Aliphatic C9-C12	8449016.000	60.000	3.111	6.377	140816.933	138488.174	-1.682
Aliphatic C12-C16	5420105.000	40.000	6.378	9.762	135502.625	137713.071	1.605
Aliphatic C16-C21	7812642.000	60.000	9.763	13.116	130210.700	135019.289	3.561
Aliphatic C21-C28	9849125.000	80.000	13.117	16.769	123114.063	127207.526	3.218
Aliphatic C28-C40	12364276.000	120.000	16.770	21.574	103035.633	103638.860	0.582
Aliphatic EPH	43895164.000	360.000	3.111	21.574	121931.011	123700.655	1.431