

Continuing Calibration Report for SequenceID : FD020725AR

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FD049063.D

Aromatic C10-C12	6706243.000	40.000	4.076	5.794	167656.075	155230.631	-8.005
Aromatic C21-C36	27654661.000	180.000	12.663	18.072	153637.006	142502.922	-7.813
Aromatic EPH	57323820.000	360.000	4.076	18.072	159232.833	149900.883	-6.225
Aromatic C12-C16	10058917.000	60.000	5.795	8.398	167648.617	162211.796	-3.352
Aromatic C16-C21	12903999.000	80.000	8.399	12.662	161299.988	154648.237	-4.301