

Continuing Calibration Report for SequenceID : FD021025AR

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FD049066.D

Aromatic C10-C12	6890054.000	40.000	4.076	5.793	172251.350	155230.631	-10.965
Aromatic C12-C16	10374039.000	60.000	5.794	8.397	172900.650	162211.796	-6.589
Aromatic C16-C21	13253164.000	80.000	8.398	12.659	165664.550	154648.237	-7.123
Aromatic C21-C36	27785097.000	180.000	12.660	18.066	154361.650	142502.922	-8.322
Aromatic EPH	58302354.000	360.000	4.076	18.066	161950.983	149900.883	-8.039