

Continuing Calibration Report for SequenceID : FD021025AR

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FD049069.D

Aromatic C10-C12	6853867.000	40.000	4.076	5.793	171346.675	155230.631	-10.382
Aromatic C12-C16	10244434.000	60.000	5.794	8.397	170740.567	162211.796	-5.258
Aromatic C16-C21	13160003.000	80.000	8.398	12.659	164500.038	154648.237	-6.370
Aromatic C21-C36	27903333.000	180.000	12.660	18.066	155018.517	142502.922	-8.783
Aromatic EPH	58161637.000	360.000	4.076	18.066	161560.103	149900.883	-7.778